G<u><u><u></u><u></u><u></u><u></u><u></u>**INSTEK**</u></u>

GPT-15000 Specifications

The specifications apply when the GPT-15000 is powered on for at least 30 minutes under +15°C~+35°C

GPT-15004 (Front)



GPT-15003/15002/15001 (Front)



GPT-15004 (Rear)



GPT-15003/15002/15001 (Rear)



(All models are available with optional GPIB or LAN)

Model \ Func.	AC Withstanding	DC Withstanding	Insulation Resistance	Ground Bond	Ground Continuity
GPT-15001	٧*				v
GPT-15002	٧*	v			v
GPT-15003	٧*	v	V		V
GPT-15004	٧*	V	V	V	v

* : Short current > 200mA

AC WITHSTANDING				
Output-Voltage Range	0.050kV~5.000kV			
Output-Voltage Resolution	1V			
Output-Voltage Accuracy	\pm (1% of setting + 5V) [no load]			
Maximum Rated Load	500 VA (5kV/100mA)			
Maximum Rated Current	100mA (0.5kV< V ≦5kV)			
	10mA (0.05kV \leq V \leq 0.5kV)			
Output-Voltage Waveform	Sine wave			
Output-Voltage Frequency	50 Hz / 60 Hz selectable			
Voltage Regulation	\pm (1% + 5V) [maximum rated load $ ightarrow$ no load]			
Voltmeter Accuracy	\pm (1% of reading + 5V)			
Current Measurement Range	1μA~100.0mA			
Current Best Resolution	1µА / 10µА / 100µА			
Current Measurement Accuracy	\pm (1.5% of reading + 30 μ A)			
Current Offset	60μA Maximum			
Window Comparator Method	Yes			
ARC Detect	Yes			
RAMP UP (Rise Time)	0.1s~999.9s			

GOOD WILL INSTRUMENT CO., LTD.

G<u><u></u>UINSTEK</u>

 N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan

 T (886) 2 2268-0389
 F (886)2 2268-0639
 www.gwinstek.com

RAMP DOWN (Fall Time)	0.0s~999.9s		
TIMER (Test Time)	OFF, 0.3s~999.9s		
TIMER Accuracy	±(100ppm + 20ms)		
WAIT TIME	0.0s~999.9s		
GND	ON/OFF		
DC WITHSTANDING			
Output-Voltage Range	0.050kV~6.000kV		
Output-Voltage Resolution	1V		
Output-Voltage Accuracy	± (1% of setting + 5V) [no load]		
Maximum Rated Load	100W (5kV/20mA)		
Maximum Rated Current	$20\text{mA} (0.5\text{kV} < \text{V} \leq 6\text{kV})$		
	$2mA (0.05kV \le V \le 0.5kV)$		
Voltage Regulation	\pm (1% + 5V) [maximum rated load \rightarrow no load]		
Voltmeter Accuracy	$\pm (1\% \text{ freading} + 5\text{V})$		
Current Measurement Range	$\pm (1\% \text{ of reading + 5V})$ $1\mu A^2 20.00 \text{mA}$		
Current Best Resolution	0.1μΑ /1μΑ /10μΑ		
Current Measurement Accuracy			
	\pm (1.5% of reading + 3 μ A) when I Reading < 1mA \pm (1.5% of reading + 30 μ A) when I Reading \ge 1mA		
Current Offset			
Window Comparator Method	5μA Maximum Yes		
ARC Detect	Yes		
RAMP UP (Rise Time)	0.1s~999.9s		
RAMP DOWN (Fall Time)	0.0s~999.9s		
TIMER (Test Time)	OFF, 0.3s~999.9s		
TIMER Accuracy	±(100ppm + 20ms)		
WAIT TIME	0.0s~999.9s		
GND	ON/OFF		
INSULATION RESISTANCE			
Output Voltage	50V~5000V dc		
Output-Voltage Resolution	50V 5000 uc		
Output-Voltage Accuracy	± (1% of setting + 5V) [no load]		
Resistance Display			
Test Voltage	Display Range		
50V≦V≦100V	0.1ΜΩ~ 10.00GΩ		
150V≦V≦450V	0.1ΜΩ~ 20.00GΩ		
150V ≦ V ≦ 450V 500V ≦ V ≦ 5000V	0.1MΩ~ 50.00GΩ		
Resistance Measurement	0.11/12 50.0002		
	Measurement Range / Accuracy		
Test Voltage 50V≦V≦450V	$0.1M\Omega^{-1}M\Omega$: ±(5% of reading + 3 count)		
$500 \ge 0 \ge 4500$	$1.1M\Omega^{50}M\Omega : \pm (5\% \text{ of reading + 3 count)}$		
	$50.1M\Omega^2 2G\Omega : \pm (10\% \text{ of reading} + 1 \text{ count})$		
500V≦V≦1200V	$0.1M\Omega^{-1}M\Omega$: ±(5% of reading + 3 count)		
	1.1MΩ~500MΩ : \pm (5% of reading + 1 count)		
	500.1MΩ~9.999GΩ : \pm (10% of reading + 1 count)		
	$10G\Omega^{5}0G\Omega : \pm (20\% \text{ of reading} + 1 \text{ count})^*$		
1250V≦V≦5000V	$0.1M\Omega^{-1}M\Omega$: ±(5% of reading + 3 count)		
	$1.1M\Omega^{500}M\Omega$: ±(5% of reading + 1 count)		
	$500.1M\Omega^{-9.999}G\Omega \div \pm (10\% \text{ of reading} + 1 \text{ count})$		
	10GΩ~50GΩ:±(15% of reading + 1 count)*		
Voltage Regulation	\pm (1% + 5V) [maximum rated load $ ightarrow$ no load]		
Voltmeter Accuracy	± (1% of reading + 5V)		
Short-Circuit Current	10mA max.		
Output Impedance	2kΩ		
Window Comparator Method	Yes		

GOOD WILL INSTRUMENT CO., LTD.

N0. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

RAMP UP (Rise Time)	0.1s~999.9s	
RAMP DOWN (Fall Time)	0.0s~999.9s	
TIMER (Test Time)	OFF, 0.3s~999.9s	
TIMER Accuracy	±(100ppm + 20ms)	
WAIT TIME	0.0s~999.9s	
GND	ON/OFF	
Ground Bond		
Output-Current	03.00A~32.00A ac	
Output-Current Resolution	0.01A	
Output-Current Accuracy	3A≦I≦8A∶±(1% of reading + 0.2A)	
	8A <i≦32a:±(1% +="" 0.05a)<="" of="" reading="" td=""></i≦32a:±(1%>	
Test-Voltage	8Vac max (open circuit)	
Test-Voltage Frequency	50Hz/60Hz selectable	
Ohmmeter Measurement Range	$1m\Omega^{\sim} 650m\Omega$	
Ohmmeter Measurement Resolution	0.1mΩ	
Ohmmeter Measurement Accuracy	$\pm(1\% \text{ of reading} + 2 \text{ m}\Omega)$	
Window Comparator Method	Yes	
TIMER (Test Time)	0.3s~999.9s	
TIMER Accuracy	±(100ppm + 20ms)	
Test Method	Four Terminal	
GND	ON/OFF	
Continuity Test		
Output-Current	100mA dc (fixed)	
Ohmmeter Measurement Range	0.10Ω~ 70.00Ω	
Ohmmeter Measurement Resolution	0.01Ω	
Ohmmeter Measurement Accuracy	$\pm(10\% \text{ of reading} + 2 \Omega)$	
Window Comparator Method	Yes	
TIMER (Test Time)	0.3s~999.9s	
TIMER Accuracy	±(100ppm + 20ms)	
MEMORY		
Single Step Memory	MANU: 100 blocks	
Automatic Testing Memory	AUTO : 100 blocks, manu per auto : 10	
INTERFACE		
REMOTE (Front) terminal	Standard	
USB host (Front)	Standard	
Rear Output	Standard	
RS-232C	Standard	
USB device	Standard	
Signal I/O	Standard	
GPIB	Option	
LAN	Option	
DISPLAY		
	7" color LCD	
POWER SOURCE		
	AC 100V~240V ± 10%, 50Hz/60Hz	
	Power consumption : 1000VA max.	
DIMENSION & WEIGHT		
GPT-15001/15002/15003	380(W) x 148(H) x 492(D) mm; Approx. 17kg	
GPT-15004	380(W) x 148(H) x 546(D) mm; Approx. 21kg	

* When Ground Mode is "ON", the measurement range is 30GΩ max. and adding 10% error for accuracy.