NO. 7-1, Jhongsing Road, Tucheng Dist., New Taipei City, 236, Taiwan

## T (886) 2 2268-0389 F (886)2 2268-0639 www.gwinstek.com

## **DAQ Modules Specifications**

Module description	Туре	Speed (ch/sec)	Max volts	Max amps	Bandwidth	Thermal offset	Comments
<b>DAQ-900</b> 20 ch Multiplexer	2-wire solid-state (4-wire selectable)	450	120V		10MHz	< 4 μV	Built-in cold junction reference
DAQ-901 20 ch Multiplexer + 2 ch current	2-wire armature (4-wire selectable)	80	300V	1A	10MHz	< 4 μV	Built-in cold junction reference 2 additional current channels (22 total)
DAQ-903 40 ch Single-Ended Mux	1-wire armature (common low)	80	300V		10MHz	< 1 μV	No four-wire measurements
<b>DAQ-904</b> 4 x 8 Matrix	2-wire armature		300V		10MHz	< 1 μV	
DAQ-908 20 ch Actuator/General Purpose Switch	SPDT/form C		300V		10MHz	< 4 μV	
DAQ-909 8 ch HV Multiplexer + 2 ch current	2-wire armature (4-wire selectable)	60	DC 600V AC 400V	2A	10MHz	< 4 μV	2 additional current channels (10 total)

## Internal DMM measurement functions supported

	DAQ-900	DAQ-901	DAQ-903	DAQ-904	DAQ-908	DAQ-909
AC/DC Voltage	√ <sup>2,3</sup>	٧	٧			٧
AC/DC Current		٧				٧
Frequency/Period	٧	٧	٧			٧
2Wire Resistance	√ ¹	٧	٧			٧
4Wire Resistance	√ ¹	٧				٧
Thermocouple	٧	٧				√ <sup>4</sup>
2Wire RTD		٧	٧			٧
4Wire RTD		٧				٧
Transistor		٧	٧			٧
Capacitance		٧	٧			٧

<sup>1.</sup> For the measurement of 100  $\Omega$  and 1 k $\Omega$  resistance ranges, it is recommended to use 4-wire resistance. The maximum resistance range of DAQ-900 is 1 M $\Omega$ .

<sup>2.</sup> When measuring AC voltage, the input impedance will decrease with frequency. A source impedance of 5  $\Omega$  or less will maintain specification over frequency. A source impedance of 50  $\Omega$  or less will maintain specification in the 5 kHz range.

<sup>3.</sup> For DC voltage measurement, if the integration time is short and the source impedance is high, more stabilization time may be required.

<sup>4.</sup> Need to use an extension cable moving the cold junction outside the chassis and manually set the reference temperature value